

Monthly radioactivity measurement "October, 2012" was conducted.**Measurement Result:**

No radioactive substances (Iodine/Caesium) was detected with any of Products/Cannulae.

Table 1. Radioactivity measurement - October/2012

Sample Category	Nuclide				Radiation Dose	
	Iodine	Caesium				
	I-131	Cs-134	Cs-137	Cs-136		
Unit	Bq/kg(L)	Bq/kg(L)	Bq/kg(L)	Bq/kg(L)	μSv/h	cpm
Products (Needle)	ND ^{*1}	ND ^{*1}	ND ^{*1}	ND ^{*1}	—	—
Cleaning Water	ND ^{*1}	ND ^{*1}	ND ^{*1}	ND ^{*1}	—	—
Cannulae	Measurement once in two months (Next November)					
Components	Measurement once in two months (Next December)					

*1) Not Detected (Detection Lower Limit 4~8)