

Radiation Measurement Plan (April 2012 ~ March 2013)

Revision : April 2012

Revising measurement frequency

We have been conducting radiation monitoring of our products and several components for one year since the earthquake, and only detected values equivalent to background or found them even lower than detection limit. Under current situation that local government has established monitoring procedures for water and air dose, we are revising our measurement planning as below.

Products	every month	➡	alternative month
Cleaning water	alternative month	➡	every three month
Cannulae/Components	alternative month	➡	every three month
Airconditioner Filter	unchanged		

* Measurement method and countermeasure at occurrence of abnormality is as described in RED below, unchanged from former procedures.

	Products	Cleaning Water	Cannulae	Components	Airconditioner Filter	Notes
	Once / 2M	Once / 3M	Once / 3M	Once / 3M	Once / 6M	
2012						
Apr.	○	○		△		
May.			△			
Jun.	○			△		
Jul.		○			○	Previous measurement in Jan. 2012
Aug.	○		△			
Sep.				△		
Oct.	○	○				
Nov.			△			
Dec.	○			△		
2013						
Jan.		○			○	
Feb.	○		△			
Mar.				△		2 years after quake

○ : Sampled at UNISIS Saitama Factory

△ : Include parts manufactured outside UNISIS

* Measurement procedure is outsourced to external inspection institute

* On detection of radiation exceeding national permissible limit, it shall be immediately reported to Government, and in parallel notify to customers and stop all shipments.